

# Interventional retrieval of a balloon-expandable stent in an infant following tetralogy of Fallot repair

Jacek Kuźma<sup>1</sup>, Michał Buczyński<sup>1</sup>, Mohamed Sameh Emam<sup>2</sup>, Wojciech Mądry<sup>1</sup>, Mariusz Kuśmierczyk<sup>1</sup>

<sup>1</sup>Department of Cardiothoracic and Transplantology, Medical University of Warsaw, Warszawa, Poland

<sup>2</sup>Student Scientific Club, Department of Cardiothoracic and Transplantology, Medical University of Warsaw, Warszawa, Poland

**Correspondence to:**

Jacek Kuźma, MD,  
Department of Cardiothoracic  
and Transplantology,  
Medical University of Warsaw,  
Żwirki i Wigury 63A, 02-091  
Warszawa, Poland,  
phone: +48 22 317 98 81,  
e-mail: jacek.kuzma@wum.edu.pl

Copyright by the Author(s), 2024

DOI: 10.33963/v.phj.101478

**Received:**

May 18, 2024

**Accepted:**

July 8, 2024

**Early publication date:**


July 10, 2024

Percutaneous interventions with pulmonary stent deployment in infants are rescue procedures for critically ill patients with poor reoperation outcomes. Stent dislodgment can lead to significant complications requiring cardiac surgery and can be a result of technical difficulties, anatomical challenges, improper stent size, or type (e.g., self-expanding stents may have a higher risk of dislodgment compared to balloon-expandable stents) [1–4].

A 10-month-old female infant was referred for tetralogy of Fallot repair. In her medical history, anoxic attacks were observed at 8 months requiring beta blocker therapy and interventional percutaneous balloon angioplasty for right ventricular outflow tract (RVOT) obstruction. Transthoracic echocardiogram, coronary angiography and computed tomography showed an additional coronary artery from the right coronary artery crossing RVOT, which made the surgery challenging. The repair was performed in cross clamp circulation. Right ventriculotomy was necessary with double patch technique (below and above coronary artery crossing RVOT) avoiding myocardial infarction. Postoperative transthoracic echocardiogram showed residual subpulmonary and proximal pulmonary branches stenosis requiring reoperation with muscular band resection and pericardial patch plasty for pulmonary arteries. Postoperative RV failure was still present due to residual outflow obstruction, pulmonary branches stenoses, and severe pulmonary and tricuspid valvular regurgitation.

The child was qualified for heart catheterization. The right femoral vein was punctured, a 6F sheath was introduced, and

anticoagulation ensured with a heparin infusion of 50 U/kg. Hemodynamic evaluation showed high central venous (16 mm Hg) and RV pressures (80% of systemic pressure) with pressure gradient across RVOT of 18 mm Hg. Right ventriculography confirmed residual subpulmonary as well as both pulmonary branches stenoses (Figure 1A–C). The child was qualified for pulmonary branches stent implantation. A Palmaz Genesis 8 mm/12 mm stent was planned to be deployed into the right pulmonary artery. However, the stent was unstable while crossing RVOT and slipped partially from the balloon catheter without any chance of removal through the 6F sheath. We considered different options for stent retrieval, including surgical cross clamp circulation, interventional forceps, a gooseneck snare loop, or stent deployment in a peripheral vein (e.g. iliac or femoral vein with a high risk of thrombosis and occlusion). The right jugular vein was punctured and a wide 9F sheath was introduced. An Amplatz gooseneck snare 7 mm loop was established over a 0.018 inch guidewire. The balloon-expandable stent was gently inserted into the 9F sheath and removed from the body (Figure 1D–E). Finally, a Palmaz Genesis 8 mm/15 mm stent was deployed successfully into the right pulmonary artery over a 0.035 inch guidewire, providing effective pulmonary flow (Figure 1F) and RV pressure drop to 60% of systemic pressure. The left pulmonary artery was bent without the chance of interventional balloon plasty or stent deployment. The general condition of the child improved and she was discharged home within 2 weeks. Reoperation with



**Figure 1.** Stent retrieval with a gooseneck snare loop in an infant following tetralogy of Fallot repair with residual subpulmonary and both branches stenosis. **A.** Right ventriculography in lateral view showing dilated right ventricle with residual subpulmonary stenosis just below localization of the coronary artery crossing the right ventricular outflow tract (white arrow). **B.** Pulmonary trunk angiography in cranial view (40 degrees) showing right (white arrow) and left (yellow arrow) pulmonary branches stenosis following surgical plasty. **C.** Right ventriculography in anteroposterior view, showing severe right pulmonary artery stenosis (white arrow). **D.** Stent partially slipped from balloon catheter over 0.018 inch guidewire. **E.** Stent inserted into 9F sheath over gooseneck snare loop. **F.** Right pulmonary artery angiography following Palmaz 8 mm/15 mm stent deployment

valved prosthesis implantation and residual muscular band resection was postponed for a few months.

Gooseneck snare devices and wide sheaths are used routinely to retrieve dislodged stents from the circulation. The presented case shows the clinical benefit of a successful procedure without the need for repeat open surgery.

### Supplementary material

Supplementary material is available at [https://journals.viamedica.pl/polish\\_heart\\_journal](https://journals.viamedica.pl/polish_heart_journal).

### Article information

**Conflict of interest:** None declared.

**Funding:** None.

**Open access:** This article is available in open access under Creative Commons Attribution-Non-Commercial-No Derivatives 4.0 International (CC BY-NC-ND 4.0) license, which allows downloading and

sharing articles with others as long as they credit the authors and the publisher, but without permission to change them in any way or use them commercially. For commercial use, please contact the journal office at [polishheartjournal@ptkardio.pl](mailto:polishheartjournal@ptkardio.pl)

### REFERENCES

1. Monello A, Biagi A, Valenti G, et al. Stent loss and successful retrieval with the loop snare technique [article in Italian]. *G Ital Cardiol (Rome)*. 2021; 22(9 Suppl 1): 435–445, doi: [10.1714/3675.36616](https://doi.org/10.1714/3675.36616), indexed in Pubmed: 34590624.
2. Lu MY, Wang KT. Coronary stent off-wire dislodgement: Case report of a rare complication. *Tex Heart Inst J*. 2023; 50(2): e207505, doi: [10.14503/THIJ-20-7505](https://doi.org/10.14503/THIJ-20-7505), indexed in Pubmed: 37023794.
3. Monteleone PP, Yeh RW. Management of complications. *Interv Cardiol Clin*. 2016; 5(2): 201–209, doi: [10.1016/j.iccl.2015.12.005](https://doi.org/10.1016/j.iccl.2015.12.005), indexed in Pubmed: 28582204.
4. Alomar ME, Michael TT, Patel VG, et al. Stent loss and retrieval during percutaneous coronary interventions: A systematic review and meta-analysis. *J Invasive Cardiol*. 2013; 25(12): 637–641, indexed in Pubmed: 24296383.