Supplementary material

Kempa M, Budrejko S, Zienciuk-Krajka A, et al. Subcutaneous implantable cardioverter-defibrillators for the prevention of sudden cardiac death: five-year single-center experience. Kardiol Pol. 2020; 78: 447-450. doi:10.33963/KP.15235

Please note that the journal is not responsible for the scientific accuracy or functionality of any supplementary material submitted by the authors. Any queries (except missing content) should be directed to the corresponding author of the article.

Table S1. Contraindications to defibrillation testing.

Patient	Contraindications to defibrillation testing
number	
9	Atrial fibrillation with inadequate anticoagulation (the test was performed later,
	after appropriate preparation of the patient with anticaogulation and
	transesophageal echocardiography to exclude left atrial thrombi)
13	Severe heart failure
20	Post-resuscitation damage of the central nervous system
23	Severe heart failure

Table S2. Cardiac electronic implantable devices (CIEDs) coexisting with subcutaneous implantable cardioverter-defibrillator (S-ICD).

Patient	Cardiac implantable electronic devices in follow-up
number	
5	Single chamber VVI pacemaker was upgraded to epicardial DDD by implantation
	of another epicardial lead onto the wall of the right atrium
6	Explantation of pacemaker can of an inactive DDD. No consent for removal of the
	leads.
8	ICD with old abandoned atrial lead after previously failed attempt of transvenous
	extraction. No further extraction attempt was undertaken.
10	Extraction of previously implanted single-chamber ICD system with a failed lead
11	Extraction of previously implanted single-chamber ICD system with a failed lead
13	Pacemaker and S-ICD removed at the time of heart transplant
14	Extraction of previously implanted single-chamber ICD system with a failed lead
15	No consent for ICD extraction
18	Extraction of previously implanted single-chamber ICD system with a failed lead