

Supplementary File

Table S1. Results of real plan (RP) *ex vivo* dosimetry on the phantom and RP and corrected plan (CP) doses

Measure point	Film dose [Gy]	SD	RP dose [Gy]	CP dose
1	1.59	0.04	1.45	1.38
2	1.84	0.02	1.60	1.66
3	1.78	0.01	1.65	1.65
4	1.40	0.02	1.48	1.48
5	1.53	0.03	1.23	1.33
6	1.69	0.03	1.60	1.56

SD — standard deviation